

# Stockport Health Authority

SUPPLIES DEPARTMENT, CHERRY TREE HOSPITAL, CHERRY TREE LANE, STOCKPORT, SK2 7PZ.

Please ask for ..... Mr. Langton .....

Your Ref: PJJ/PAM

Our Ref:

061.419.8400  
Telephone 061-488-8816/7

13th May, 1987

Vianed,  
15 Station Road  
Cross Hills  
Keighley,  
West Yorkshire,  
BD20 7DT


For the attention of Mr. J. S. Lamb

Dear Sirs,

You will recall that in July 1985 I notified you of what this Authority considered to be a potential hazard with regard to cot tops supplied by yourself, and in fact, following investigations by both the D.H.S.S./S.T.B. and yourselves modifications were carried out with regard to the cot tops.

I have now received information from the Maternity Unit at Stepping Hill Hospital that in the period of the last 7 weeks a further two incidents have been reported where on both occasions babies have kicked the cot tops which have then fallen into the cot. Fortunately on both occasions no injuries occurred to the children. Bearing in the mind the potential dangers which could again occur, I shall be grateful if your representative would contact me so that he can arrange a visit to the site and we can investigate this matter with some urgency.

Yours faithfully,



Customer Services Manager

- 1) JSL spoke on tel. and matter not to go further
- 2) SH went in and is not so sure the unit are satisfied GGL

in pencil on original in Stepping Hill file