

RE: SRS69177

1 message

Stephen Machin <stephen@prosol.co.uk>
To: Cathy Green <cathy.green@viamed.co.uk>

21 July 2025 at 15:47

Hi Cathy,

Apologies for the delayed response, I have been away on holiday.

Please send back the sensors with no faults with one of our next orders to save postage charges.

Please dispose of the out of warranty sensor.

Thanks

Best Regards

Stephen Machin | Operations Manager | Prosol UK Sales & Distribution Ltd T: 0114 255 7700 | purchases@prosol.co.uk | www.prosol.co.uk





From: Cathy Green <cathy.green@viamed.co.uk>

Sent: 08 July 2025 17:22

To: Stephen Machin <stephen@prosol.co.uk>

Subject: SRS69177

EXTERNAL SENDER WARNING. This message was sent from outside your organisation. Please do not click links or open attachments unless you recognise the source of this email and know the content is safe.

Dear Mr Machin

We have received and tested the oxygen sensor that were returned to us on SRS69177, please see results below.

Sensors under warranty.

P/N:0110365 S/N: P106590 No Fault Found

P/N:0110122 S/N: 218618 No Fault Found

Sensor out of warranty P/N:0110122 S/N:793096 Sensor has no out put. Please advise how you would like to proceed. Thank you in advance. Kind regards Cathy Green Goods Out / Stock Control Supervisor Viamed Ltd. http://www.viamed.co.uk Email: cathy.green@viamed.co.uk Tel: 44 (0)1535 634 542 Fax: 44 (0)1535 635 582 Twitter: twitter.com/ViamedLtd Facebook Page: Search for Viamed Ltd Company registered in England, No. 01291765. E-mail Disclaimer GDPR / Privacy Policy available at www.viamed.co.uk/policy The information in this email is confidential and may be legally privileged. It is intended solely for the addressee. Access to this email by anyone else is unauthorized. If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful. If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call 44 (0)1535 634542