

Service Repair Sheet 69160
Contact Name CM Hor
Company/ Hospital Name Emstec Services
Direct Phone 604 658 1791
General Phone 6046581791
Opera Account 00013104
Email cmhor@emstecmedical.com
Date Received 23/Jun/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37961	1410000	PR02041A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
