

Service Repair Sheet SRS69194

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Kim Wareham

Gemco UK LTD

Spares Sales Administrator

0345 872 2400

03458722400

00020099

Kim.Warham@gemco.co.uk

22/Jul/2025

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____
Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 18/Jul/2025 Sophie Lines

18/Jul/2025 Sophie Lines

Customer has a R-22AVG giving faulty readings - serial number V118077. Should be under warranty

22/Jul/2025 Robert Connor

Received 1 x R22AVG s/n V118077.

Ready For quote

Repair Complete Signed

C Green 30.7.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38013	Oxygen Sensor	0110182	V118077	Yes

0140013 x 1 FOC
old S/N, SRS, SRN

0110182 x 1 FOC
new S/N, SRS, SRN
warranty replacement

UPS x 1 £12.