

Gender	Age	PMH	Comments
M	3	Respiratory Chest Issues	02:00 - woke up crying as did not want it on foot, refusing to have sensor back on, 04:30am - still refusing to have sensor on and kicking it off, does not want the machine on his toe
M	10	Duchennes Muscular Dystrophy	not the best night's sleep due to the wires and the machine buzzing
M	1	Down Syndrome	probe detached from toe after 1.5 hours
M	14	Epilepsy	2 nights - probe off 3 times first night, 2 times second night
M	0.88	Parental reports of breathing pauses	Night 1 - 10:15 woke after 2 hours as monitor off foot. 11:45pm - finally reattached to toe; 03:15am - turned off machine as patient woke up and machine fell off foot. Would not let me put it back on. Night 2 - 09:30pm attached; 09:44pm woke up - put it back on as wasn't on toe properly; 12:30 - turned off as patient fully awake and won't let parent put it back on. 01:00 am reattachedm working ok. 01:45am - turned off for the night as won't keep it on foot and distressed
F	3	Sleep related breathing disorder	Night 1 - 20:42 bedtime; 20:47 - asleep; 01:15 - came off toe ?exact time; 07:52- awake Night 2 - 18:30-asleep; 20:33 - came off toe; 07:15 - awake
F	3	Wheeze	Overnight - woke up several times taking off the wire and falling back to sleep. Had to reattach the wire using up most of the stickers
F	2	Choking, ?asthma	Moved wire and probe causing machine to alarm. Probe kept falling off causing machine to alarm.
M	7	?sleep disordered breathing. Autism	Tried with machine at home for 2 nights - failed, no data