

Service Repair Sheet 69184

Contact Name	Moynul Rashid
Company/ Hospital Name	Queens Hospital/Siemens Medical Solutions
Position	Siemens MES Technical Administrator
Direct Phone	01708435482
General Phone	01708435482
Opera Account	00000264
Email	moynul.rashid@siemens-healthineers.com
Order Number	4532636829
Date Received	09/Jul/2025
Booked in By	Kate Griffiths

Repair	Ref	S/N	Equipment Type	
SRN38003	1410000	PR02912A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
