

Service Repair Sheet 69206

Contact Name	Hocine Benabdelhafid
Company/ Hospital Name	Homerton Hospital
Department	Medical Electronics & Physics
Position	Calibration Officer
Direct Phone	020 8510 7935
General Phone	02085105555
Opera Account	00002950
Email	hocine.benabdelhafid@nhs.net
Order Number	334126472
Date Received	12/Aug/2025
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38100	1410000	PR01702A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.

---

Repair	Ref	S/N	Equipment Type	
SRN38101	1410000	PR02260A14	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.

---