



Kate Griffiths <viamed.kate.griffiths@gmail.com>

RE: MORGATE Biomédical/ Order PO2503-00102

1 message

MORGATE info <info@morgate.fr>

1 September 2025 at 14:00

To: Kate Griffiths <kate.griffiths@viamed.co.uk>

Dear Kate,

Our customer comes to us about MX300 SN 340510. It displays “erreur 71” when starts.

The user has replaced batteries, O2 sensor, O2 cable but still the same problem.

Could you tell me if it can be repaired?

Thank you in advance for your information.

Best Regards / Bien Cordialement

Karine BOUCHER

Assistante

MORGATE Biomédical

23 Av. de L'Europe, Le Newton, Batiment 2, 78400 CHATOU

Tél. : 09 72 64 34 96 Fax : 09 72 63 77 48 info@morgate.fr www.morgate.fr

