

# Service Repair Sheet SRS69215

Contact Name  
Company/ Hospital Name  
Department  
Position  
Direct Phone  
General Phone  
Opera Account  
Email  
Order Number  
Date Received  
Booked in By  
Main Company  
Type Return

Joanne Van De Geer  
Delta Medic BV

0031 30 274 1302

0031302741302

CID28200

info@deltamedic.nl

DVM156489-1

27/Aug/2025

Robert Connor

Viamed

Warranty

## VIAMEDclean

### Goods In Only

Decontamination  
certificate provided  
by customer

Cleaned by Viamed,  
if no declaration certificate  
from customer

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

### Goods Out Only

Cleaned by Viamed before  
returning to customer

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

Notes 18/Aug/2025 Aqib Majeed

18/Aug/2025 Aqib Majeed

Customer returning 2 x r-30v sensors SN: V108431 and V108433 as they have a gas leak. still under warranty.

27/Aug/2025 Robert Connor

Received 2 x R30V s/n V108431, V108433.

Ready For quote

Repair Complete Signed

CGreen 1-9-25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38141	Oxygen Sensor	0110040	V108431	Yes
SRN38142	Oxygen Sensor	0110040	V108433	Yes

0140013 x 2 FOC  
Old SIN, SRS, SRN

0110040 x 2 FOC  
New SIN, SRS, SRN

UPS x 1 FOC.