

EU Declaration of Conformity

This declaration of conformity is issued under the sole responsibility of the manufacturer in accordance with Regulation 2017/745/EU Article 120 in its currently valid version. Article 120 (3a) case b) applies to the products. The requirements from Article 120 (3c) are met.

Product: **Oxygen Sensors**
Basic UDI-DI: **4036616OOMCK**

Type:	Honeywell	Viamed	Honeywell	Viamed	Honeywell	Viamed
	OOM101	R15V	OOM105	T-7V	OOM201	R-23V
	OOM102	R-22Vi	OOM106	R-45V	OOM202	R-49V
	OOM102-A	R-22VA	OOM110	R-30V	OOM202-1	R-41V
	OOM102-1	R-17Vi	OOM111	R-45V	OOM202-2	R-44V
	OOM103	R-22MEDV	OOM112	R-43V	OOM202-2S	R-42V
	OOM103-1	R-17V	OOM113	R-48V	OOM204	R-47V
	OOM104	R-23Vi				

Classification: **Class IIa**
(Annex VIII)

CE Mark: 

Notified Body: TÜV SÜD Product Service GmbH, Ridlerstr. 65, 80339 München, Germany

Conformity Assessment Process: Annex II, section 3 of Directive 93/42/EEC

Particular product standard applied: ISO 80601-2-55

Issued by: **Honeywell Healthcare Solutions GmbH**
Eudamed-SRN: DE-MF-000004924
Alter Holzhafen 18
D-23966 Wismar
Germany

Valid from: 04.03.2024

Valid until: 31.12.2028

Place, Date: Wismar, 04.03.2024

Authorized Signature:


ppa. Marko Sprössel


ppa. Steffen Wickner

Honeywell Healthcare Solutions GmbH
Alter Holzhafen 18
23966 Wismar / GERMANY
Tel.: +49(0)3841 / 360 - 1
Fax: +49(0)3841 / 360 - 222