

EU Declaration of Conformity

This declaration of conformity is issued under the sole responsibility of the manufacturer in accordance with Regulation 2017/745/EU Article 120 in its currently valid version. Article 120 (3a) case b) applies to the products. The requirements from Article 120 (3c) are met.

Product: **Oxygen Sensors**
Basic UDI-DI: **4036616OOMCK**
Type: OOMLF102 OOMLF103 OOMLF111
OOMLF102-1 OOMLF103-1 OOMLF202
OOMLF102-L OOMLF103-1M OOMLF202-1
OOMLF102-1L OOMLF106 OOMLF204
OOMLF102-M3 OOMLF110

Classification: **Class IIa**
(Annex VIII)

CE Mark: 

Notified Body: TÜV SÜD Product Service GmbH, Ridlerstr. 65, 80339 München, Germany

Conformity Assessment Process: Annex II, section 3 of Directive 93/42/EEC

Particular product standard applied: ISO 80601-2-55


Issued by: **Honeywell Healthcare Solutions GmbH**
Eudamed-SRN: DE-MF-000004924
Alter Holzhafen 18
D-23966 Wismar
Germany

Valid from: 04.03.2024
Valid until: 31.12.2028

Place, Date: Wismar, 04.03.2024

Authorized
Signature:


ppa. Marko Sprössel


ppa. Steffen Wickner

Honeywell Healthcare Solutions GmbH
Alter Holzhafen 18
23966 Wismar / GERMANY
Tel.: +49(0)3841 / 360 - 1
Fax: +49(0)3841 / 360 - 222