

## Calibration Certificate

The VN202 Oxygen analyser has been tested for accuracy by using a UKAS certified voltage source.

|                                    |            |
|------------------------------------|------------|
| Calibrated voltage source          | Microcal   |
| ISO calibration index number       | CE076      |
| Serial number of the Microcal      | 6939K6     |
| Certificate number of the Microcal | 1895708    |
| Serial number of the VN202         | 03847      |
| Certificate number of the VN202    | 1757428393 |

| Voltage input: | Target reading: | VN202 reading: |
|----------------|-----------------|----------------|
| 5.0mV          | 10%             | 9.8%           |
| 10.45mV        | 20.9%           | 20.9%          |
| 10.5mV         | 21%             | 21.0%          |
| 12.5mV         | 25%             | 25.0%          |
| 25.0mV         | 50%             | 50.0%          |
| 37.5 mV        | 75%             | 75.0%          |
| 50.0mV         | 100%            | 100%           |

Tested by: M.Green

Date: 09/09/2025

Signed: