

Service Repair Sheet 69209

Contact Name	Umer Tariq
Company/ Hospital Name	Umer Tariq
Direct Phone	07535 480529
General Phone	07535480529
Opera Account	CID23684
Email	umertariq616@gmail.com
Date Received	01/Sep/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38143	1410000	PR03507A14	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
