

Service Repair Sheet 69180
Contact Name Barry Moore
Company/ Hospital Name Frimley Park Hospital
Department Clinical Engineering Department
Direct Phone 03006136332
General Phone 01344877482
Opera Account 00001740
Email barry.moore@nhs.net
Order Number P785337
Date Received 15/Sep/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38167	1410000	PR03164A10	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
