

Service Repair Sheet 69204

Contact Name	Amy Phillips
Company/ Hospital Name	Midland Metropolitan University Hospital
Direct Phone	01215074070
General Phone	01215531831
Opera Account	CID31628
Email	amy.phillips12@nhs.net
Order Number	SWBH178887
Date Received	11/Aug/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38099	1410000	PR01780A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
