

Service Repair Sheet SRS69193

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Manu Karumalil John
Darcy Healthcare

07404138177
03308189300
CID33497
manu.john@darcyhealthcare.com

27/Aug/2025
Robert Connor
Viamed
Quote

VIAMEDclean

Goods In Only
Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: RC
Date: 27/8/25

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 17/Jul/2025 Aqib Majeed

17/Jul/2025 Aqib Majeed

Customer is returning a MaxO2 Me Monitor for a Service, S/N GC48099014

27/Aug/2025 Robert Connor

Received 1 x MAXO2ME s/n GC48099014, with cable ID2687040, MAX-550E s/n GC25399015, and blue tee. Battery compartment not checked, can't unscrew.

Ready For quote

CGreen 2-9-25

Repair Complete Signed

CGreen 19-9-25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38138	Oxygen Monitor	0111263	GC48099014	2
SRN38139	Misc. Cable	0131700	Tracking5	2
SRN38140	Oxygen Sensor	0110452	GC25399015	2

0180050 x 1 @ £46.20
S/N, SRS, SRN

0140025 x 1 FOC
SRS, SRN

0140012 x 1 FOC
Old S/N, SRS

0110452 x 1 @ £112.10
New S/N, SRS, SRN

UPS x 1 @ £12