

Service Repair Sheet 69236  
Contact Name Callum Baker  
Company/ Hospital Name Chesterfield Royal Hospital  
Direct Phone 01246277271  
General Phone 01246277271  
Opera Account 00001170  
Email callum.baker@nhs.net  
Order Number 80115059  
Date Received 11/Sep/2025  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38161	1410000	PR0408A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.

---