

Service Repair Sheet SRS69243

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

David Edmondson
Royal Preston Hospital
Medical Engineering
Technician
01772 524598
01772716565
00004260
david.edmondson@lthtr.nhs.uk

22/Sep/2025
Robert Connor
Viamed
Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☐

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 17/Sep/2025 Kate Griffiths

17/Sep/2025 Kate Griffiths

They have a faulty Tom Thumb, no pressures s/n: 0401172

17/Sep/2025 Kate Griffiths

UPDATE: Unit has been sent, I did ask for cert of decont/paperwork with SRS number on

22/Sep/2025 Robert Connor

Received 1 x tom thumb s/n 0401172, pressure gauge is loose.

Ready For quote

C Green 24-9-25

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38178	Tom Thumb	0310030	0401172	N

0340000 x 0.25 @ £95 per hr = £23.75
SRS, SRN

0330223 x 1 @ £6.20
SRS, SRN

UPS x 1 @ £12