

Oxygen sensor R-22AV Acceptance inspection record

	No.	S/N	atmospheric power [mV] 9 ~ 13.5	0% oxygen output [mV] ≤0.4	vibration test [mV] ≤±0.2	judgement
K4	1	212389	14.08	0.01	0.01	NG
K4	2	212398	14.73	0.01	0.01	NG
K4	3	212405	13.78	0.01	0.01	NG
L4	4	222182	13.72	0.02	0.02	NG
BS	5	243145	0.05	0.01	0.01	NG
DS	6	257497	4.20	0.01	0.01	NG