

Service Repair Sheet 69239

Contact Name	David Watson
Company/ Hospital Name	Darlington Memorial Hospital
Department	Clinical Engineering Department
Position	Clinical Engineering Technician
Direct Phone	01325743048
General Phone	01325380100
Opera Account	00001350
Email	david.watson28@nhs.net
Order Number	CDD0006423762
Date Received	16/Sep/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38171	0310030	VMBC18	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		

Tom Thumb TT480 VMBC18 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.
