

Service Repair Sheet 69254

Contact Name	Ferdinand De Guzman
Company/ Hospital Name	University College Hospital
Department	Medical Physics and Biomedical Engineering
Position	Biomedical Engineer
Direct Phone	02034476255
General Phone	0845 1555000/02034 567890
Opera Account	00003420
Email	f.deguzman@nhs.net
Order Number	RRVN400397322
Date Received	01/Oct/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38198	1410000	PR02987A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
