

Service Repair Sheet 69246

Contact Name	Mamata Naik
Company/ Hospital Name	Glasgow Royal Maternity Hospital
Department	Medical Physics
Position	Medical Physics Technician
Direct Phone	01412 115239
General Phone	0141 211 5400
Opera Account	00001830
Email	mamata.naik@ggc.scot.nhs.uk
Order Number	GMPR15819800
Date Received	24/Sep/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38179	1410000	PR02724A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
