

★🕒 **Ryan Swaine** 4 Issues **1 Unread** 0 ISO Tasks 19.60 °C

Vasterorts Dyk AB

🔧 CID5206 Opera **V002272** Type: **Vandagraph Trade**
Vasterorts Dyk AB Sundbyvagen 83, Spange, , Sweden,
16344, Sweden
Tel **0046 8366688** GBP Trade Price List Vandagraph
(0009) [Log Call](#) [Add Issue](#)
Credit 30 Days after date of invoice Credit Limit 0.00GBP
VAT SE556604342701

[Account Home](#)[Main Details](#)[Account Memos](#)[Sales History](#)[Stock References](#)[Cust./Supp. Returns](#)[Service Visits](#)[Distributor Agreements](#)[Documents Files](#)[Linked Issues](#)[Regulatory Requirements](#)

Updating Host of Repair

[Finish Update to SRS69273](#)Quick Search: [Find / Refresh](#)

SRN Number 69273

[< ATTACH CUSTOMER LETTER\(S\) >](#)

Company / Hospital	Vasterorts Dyk AB
Contact Name	Conny Billquist
Department	
Position	Accounts
Company Phone	0046 8366688
Direct Phone	468366688
Opera Account	V002272
Email	info@nitrox.se
Order Number	
Opera Quote	QUO
Type	Warranty
Notes	22/Oct/2025 Ryan Swaine 22/Oct/2025 Ryan Swaine

Add Individual Items To SRN

Stock Reference Serialnumber Barcode ID Description Extra Track Numbers Invoice No.

Equipment Type

Unknown Date Originally Invoiced
[Date finder](#)

01/Jan/1970

Estimate warranty Date

Previous SRS Previous SRN New Warranty ☐Repair Warranty ☐

Internal Clean Procedure

Cleaned Status

NEEDS Incoming Cleaning Procedure Use
VM3COP27.51 Incoming / Goods in Contamination Control

Nothing Done

TEK-OX TX 5366 and TX 5365 showing an E4 error. To be replaced FOC. Devices are to be used for evaluation of E4 Error. 22/Oct/2025 Ryan Swaine Replaced in advice on CAN159735	Notes Specific to this Item Add To SRS 69273
---	--

Items for Repair

Unique ID	Equipment	Stock_ref	Serialnumber	Description	Warranty New	Warranty Repair
-----------	-----------	-----------	--------------	-------------	--------------	-----------------