

Service Repair Sheet 69263
Contact Name Louise Taylor
Company/ Hospital Name Royal Stoke University Hospital
Direct Phone 01782 672 338
General Phone 01782715444
Opera Account 00012132
Email clin.tech@uhnms.nhs.uk
Order Number X031848
Date Received 13/Oct/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38211	1410000	PR0466A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
