

Service Repair Sheet 69229

Contact Name	Claire Willis
Company/ Hospital Name	Medway Maritime Hospital
Department	Clinical Engineering
Position	Clinical Engineering Inventory and Systems Assistant
Direct Phone	01634976451
General Phone	01634830000
Opera Account	00001770
Email	c.willis2@nhs.net
Order Number	260008329
Date Received	10/Sep/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38160	1410000	PR0773A31	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
