

Service Repair Sheet SRS69237

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Tracy Wang

Justec Shenzhen Co. Ltd

008675526068930

008675526068931

00006027

tracy.wang@justec.cn

22/Oct/2025

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☐

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 09/Sep/2025 Ryan Swaine

09/Sep/2025 Ryan Swaine

R-23V sn. V125446, not calibrating on Draeger Evita 4.

22/Oct/2025 Robert Connor

Received 1 x R23v s/n V125446, hidden in a box of Vandagraph stock, not discovered for 3 weeks.

Ready For quote

Repair Complete Signed

C. GREEN 4-11-25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38230	Oxygen Sensor	0110023	V125446	Yes

0140013 x 1 FOC

S/N, SRS, SRN

0110023 x 1 FOC

new S/N, SRS, SRN

UPS x 1 Ex Saver
Ship with next order