

Service Repair Sheet 69282

| | |
|------------------------|-------------------------------|
| Contact Name | Mandy Marquis |
| Company/ Hospital Name | Princess Elizabeth Hospital |
| Department | EBME Department |
| Position | EBME Administrative Assistant |
| Direct Phone | 01481 707334 |
| General Phone | 01481725241 |
| Opera Account | 00004825 |
| Email | Mandy.Marquis@gov.gg |
| Order Number | 4500521462/1 |
| Date Received | 03/Nov/2025 |
| Booked in By | Robert Connor |

| | | | | |
|----------------|---------|---------|----------------|--|
| Repair | Ref | S/N | Equipment Type | |
| SRN38291 | 0310030 | 0401051 | Tom Thumb | Complete - Repaired Time :0 Hour(s) |
| Parts Replaced | | | | |
| | Qty: 1 | 0380000 | | |

Tom Thumb TT480 0401051 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.
