Service Repair Sheet 69260

Contact Name Neo Grant

Company/ Hospital Name Luton & Dunstable Hospital Department Clinical Engineering Department

Position Clinical Engineer
Direct Phone 01582497032
General Phone 01582497042
Opera Account 00003490

Email Neo.Grant@bedsft.nhs.uk

Order Number 30169833
Date Received 13/Oct/2025
Booked in By Robert Connor

Repair Ref S/N Equipment Type

SRN38210 1410000 PR02041A12 Foetal Simulator Complete - Repaired

Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000 Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.