

Service Repair Sheet 69283  
Contact Name Mark Moore  
Company/ Hospital Name St Bartholomews Hospital  
Department Clinical Engineering Department  
Position Clinical Technologist  
Direct Phone 02034656739  
General Phone 02074804622  
Opera Account 00003204  
Email mark.fabrosmoore@nhs.net  
Order Number 41212825  
Date Received 06/Nov/2025  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38296	1410000	PR02940A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.

---