

History

Sophie Lines	SRS69300	25/11/2025	Jake Jones	SRS69300 25/Nov/2025 Sophie Lines 25/Nov/2025 Sophie Lines In March 2025 we sent 4x R-47V sensors on RVM155855. Jake has emailed today highlighting an issue with their last sensor from this order - the R-47V blister pack showing serial number V107762 and showing a manufacture date of 2025-03-01 actually contained inside it an OOMLF204 sensor/serial number B110191 - an error made by Honeywell. HL and CG advised to process a FOC order to replace the incorrect sensor, and also provide an SRS number to return the incorrect cell along with all packaging. We will be providing a UPS label for the return.
--------------	-----------------	------------	------------	---