

Service Repair Sheet 69284  
Contact Name Ross Bowden  
Company/ Hospital Name Cardiff and Vale UHB  
Department Clinical Engineering  
Direct Phone 02921845678  
General Phone 02920903794  
Opera Account 00000092  
Email Ross.Bowden@wales.nhs.uk  
Order Number 726819666  
Date Received 04/Nov/2025  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38293	1410000	PR02450A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.

---