

Service Repair Sheet 69296

Contact Name	Tim Bray
Company/ Hospital Name	Medisafe Ltd
Direct Phone	01962762098
General Phone	01962885777
Opera Account	00011608
Email	tim.bray@medisafe.org.uk
Date Received	19/Nov/2025
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38359	1410000	PR02586A14	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.

---