

Service Repair Sheet SRS69293

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Nigel Baldwin

St Woolos Hospital

Medical Electronics

01633234234

01633234234

00003836

Nigel.baldwin@wales.nhs.uk

25/Nov/2025

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☐

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 18/Nov/2025 Kate Griffiths

18/Nov/2025 Kate Griffiths

rang to say 0014650 finger sensor s/n ESHK00895 does not work at all with their device and would like a replacement.

25/Nov/2025 Robert Connor

Received 1 x SPO2 sensor s/n ESHK00895

Ready For quote

Repair Complete Signed

CC Neen 27-11-25

HL agreed
to straight
swap

cg 27-11-25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38361	Pulse Oximetry Finger Sensor	0014650	ESHK00895	Yes

004001 x 1

S/N, SRS, SRN

No fault found.

Ordered wrong type of sensor
for Monitor, see line below
for replacement.

0014651 x 1 FOC

S/N, SRS, SRN

UPS x 1 FOC