



Viamed Limited · 15 Station Road · Cross Hills
 Keighley · West Yorkshire · BD20 7DT · United Kingdom
 Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582
 Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 69302

Contact Name Neo Grant
 Company/ Hospital Name Luton & Dunstable Hospital
 Department Clinical Engineering Department
 Position Clinical Engineer
 Direct Phone 01582497032
 General Phone 01582497042
 Opera Account 00003490
 Email Neo.Grant@bedsft.nhs.uk
 Order Number 30172517
 Date Received 02/Dec/2025
 Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38463	1410000	GE71743928	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.