

Service Repair Sheet SRS69311

Contact Name Fiona Halls
 Company/ Hospital Name The Automation Partnership
 Department
 Position
 Direct Phone 01763227248
 General Phone 01763227248
 Opera Account 00009081
 Email Fiona.Halls@sartorius.com
 Order Number
 Date Received 15/Dec/2025
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Warranty

VIAMEDclean	
Goods In Only	
Decontamination certificate provided by customer	
<input checked="" type="checkbox"/>	<input type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer	
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 08/Dec/2025 Kate Griffiths

08/Dec/2025 Kate Griffiths

Will send in 2 x 0110432, s/n LE90799101 and LC77899195 - will not calibrate during set up. They say they expect a replacement within 5 working days (see attached). Please see attachment for their list of expectations... Their refs: S-0262 and Defect ticket 9265

15/Dec/2025 Robert Connor

Received 2 x MAX-250(B) s/n LC77899195, LE90799101

Ready For quote

Repair Complete Signed

Caron 17.12.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38472	Oxygen Sensor	0110432	LC77899195	Yes
SRN38473	Oxygen Sensor	0110432	LE90799101	Yes

0140013 x 2 FOC

S/N's, SRS, SRN

0110432 x 2 FOC

New S/N's, SRS, SRN

UPS x 1 FOC