

Service Repair Sheet 69309

Contact Name Denis Bratley
Company/ Hospital Name Worthing Hospital
Department Clinical Engineering Department
Position Clinical Engineering Specialist
Direct Phone 01243831643
General Phone 01903205111
Opera Account 00012604
Email uhsussex.ebme.srh@nhs.net
Order Number 3392477
Date Received 15/Dec/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38471	1410000	PR0773A30	Foetal Simulator	Complete - Repaired
				Time :0 Hour(s)
Parts Replaced				
Qty: 1 1480000				
Qty: 1 1430309				

V1000 has had transducer interface cushion replaced and calibration has been checked.
