

Certificate of Calibration

Certificate number: 1767707287

Model Number: MaxBlend 2

Serial Number: FA56833001

Equipment used: Sealey STS104 Torque Screwdriver (CE222)

AX300 Oxygen Monitor (CE198)

ITI Pressure Gauge (CE215 + CE221)

Calibration traceable to UKAS calibration standards.

The output values were measured in accordance with the MaxBlend 2 Service Manual.

Dial Setting	Oxygen/Air Pressure	Oxygen Concentration Target Value	Oxygen Concentration Actual Value
21	50psi/50psi	20.9 to 22.9%	21.1
30	50psi/50psi	27.0 to 33.0%	28.5
60	50psi/50psi	57.0 to 63.0%	59.6
90	50psi/50psi	87.0 to 93.0%	88.3
100	50psi/50psi	98.0 to 100%	99.8
30	50psi/40psi	27.0 to 33.0%	28.3
60	50psi/40psi	57.0 to 63.0%	59.7
90	50psi/40psi	87.0 to 93.0%	88.3
30	50psi/60psi	27.0 to 33.0%	28.7
60	50psi/60psi	57.0 to 63.0%	60.2
90	50psi/60psi	87.0 to 93.0%	88.6
Check Duckbills			Pass
Check for leaks			Pass
Check alarm for low oxygen			Pass
Check alarm for low air			Pass

Name: Phil Crossley

Date: 6th January 2026

Signed: