

Service Repair Sheet 69308

Contact Name Michael Carrick
Company/ Hospital Name Royal London Hospital
Position Senior Clinical Technologist
Direct Phone 02035942240
General Phone 02073777000
Opera Account 00003030
Email Michael.carrick@nhs.net
Order Number 41226574
Date Received 09/Dec/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38468	0310265	TE00000023	Blender	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380061
Qty: 1	10003V2
Qty: 1	R229P06

Blender TE00000023 has been stripped and cleaned. All O rings, diaphragms etc have been replaced and it has been calibrated.
