

Service Repair Sheet 69320

Contact Name	Arlene Gallagher
Company/ Hospital Name	Forth Valley Royal Hospital
Department	Medical Physics Dept
Position	Ms Gallagher
Direct Phone	01324 566953
General Phone	01324 566000
Opera Account	00004880
Email	arlene.gallagher@nhs.scot
Order Number	VMF16157639
Date Received	30/Dec/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38485	1410000	PR01985A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
