

Service Repair Sheet 69306

Contact Name Kamal Pacitti
Company/ Hospital Name North Middlesex University Hospital
Department Clinical Engineering Department
Position Specialist Clinical Engineer
Direct Phone 02088873790
General Phone 02088872000
Opera Account 00003070
Email kamal.pacitti@nhs.net
Order Number RFL679613
Date Received 09/Dec/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38464	0310200	B1305793	Blender	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
Qty: 1 0380019				
Qty: 1 0332001				

Blender B1305793 has been stripped and cleaned. All O rings, diaphragms etc have been replaced and it has been calibrated.

Repair	Ref	S/N	Equipment Type	
SRN38465	0310200	B1305801	Blender	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
Qty: 1 0380019				
Qty: 1 0332001				

Blender B1305801 has been stripped and cleaned. All O rings, diaphragms etc have been replaced and it has been calibrated.

Repair	Ref	S/N	Equipment Type	
SRN38466	0310200	B1307007	Blender	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
Qty: 1 0380019				
Qty: 1 0332001				
Qty: 1 0332017				

Blender B1307007 has been stripped and cleaned. All O rings, diaphragms etc have been replaced and it has been calibrated.

The smashed control dial has been replaced.
