

Service Repair Sheet 69326

Contact Name Michelle Bowles
Company/ Hospital Name Norfolk and Norwich University Hospitals NHSFT
Department Clinical Engineering
Position Medical Equipment Technician
Direct Phone 01603286568
General Phone 01603286286
Opera Account 00003890
Email michelle.bowles@nnuh.nhs.uk
Date Received 12/Jan/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN38490	1410000	PR03731A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
