Certificate of Calibration

Certificate number: 1769593984

V1000 Foetal Heart Simulator
Part Number: 1410000
Serial number: PR0773A29

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with procedure: VM3COP55.01.

Equipment used: Sonicade FM6 foetal monitor (CE051)

TTi TF930 3GHz counter (CE185/CE216) - calibration traceable to UKAS calibration standards.

V1000 BPM setting	Period values stated in milliseconds				Result PASS/FAIL
	Target	Lower limit	Upper limit	Actual	
60	1000.00	990.00	1010.00	**992.70**	PASS
90	666.70	660.03	673.37	**665.37**	PASS
120	500.00	495.00	505.00	**499.82**	PASS
150	400.00	396.00	404.00	**399.96**	PASS
180	333.30	329.97	336.63	**332.50**	PASS
210	285.70	282.84	288.25	**285.71**	PASS

Name: Phil Crossley

Date: 28th January 2026

Until 27th January 2027

Signed: _____