

Service Repair Sheet 69333

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Rob Ellis

University Hospitals Plymouth NHST

Clinical Engineering

Senior Clinical Engineer

01752 431333

01752202082

00004127

rob.ellis4@nhs.net

46974048

13/Jan/2026

Robert Connor

| Repair | Ref | S/N | Equipment Type | Complete - Repaired |
|----------|---------|----------|----------------|---------------------|
| SRN38491 | 2510000 | H0001465 | Microstim DB3 | Time :0 Hour(s) |

Parts Replaced

Qty: 1 2540010

Qty: 1 2510000

Qty: 1 2520000

3 included as part of repair
Cg 28-1-26

The Microstim DB3 has been repaired and the patient lead has been replaced.
