

Service Repair Sheet 69307

Contact Name Tim Holland
Company/ Hospital Name North Manchester General Hospital
Department Clinical Engineering Dept.
Position Senior Clinical Engineer
Direct Phone 01617202347
General Phone 01617954567
Opera Account 00003580
Email timothy.holland@mft.nhs.uk
Order Number 000101_100006678
Date Received 22/Dec/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38480	0310030	040463	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 040463 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN38481	0310030	040464	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 040464 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN38482	0310030	040465	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 040465 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN38483	0310030	040466	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 040466 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.
