

Service Repair Sheet 69331

Contact Name Diane Cobbledick
Company/ Hospital Name Wansbeck General Hospital
Department Medical Electronics
Position Senior Medical Electronics Technician
Direct Phone 01670529020
General Phone 01670521212
Opera Account 00000161
Email Diane.Cobbledick@northumbria-healthcare.nhs.uk
Order Number RTFE400473564
Date Received 15/Jan/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38503	1410000	PR03124A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
