

Service Repair Sheet 69330

Contact Name Josh Cowen
Company/ Hospital Name Worthing Hospital
Direct Phone 01903205111
General Phone 01903 20 51 11
Opera Account 00004621
Email josh.cowen@nhs.net
Order Number 3403123
Date Received 22/Jan/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38512	1410000	PR0773A29	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
