

Service Repair Sheet 69347

Contact Name Janet Driver  
Company/ Hospital Name Barnet Hospital  
Direct Phone 02082164978  
General Phone 02082164978  
Opera Account 00000261  
Email janet.driver@siemens-healthineers.com  
Order Number 4532757060-A  
Date Received 28/Jan/2026  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38517	1410000	PR02315A14	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.

---