

Service Repair Sheet SRS69321

Contact Name	Damien Mahendrarajah
Company/ Hospital Name	Royal Brompton and Harefield
Department	Clinical Engineering
Position	Clinical Engineering Technologist
Direct Phone	02073528121
General Phone	020 7352 8121
Opera Account	00002790
Email	damien.mahendrarajah@nhs.net
Order Number	
Date Received	08/Jan/2026
Booked in By	Robert Connor
Main Company	Viamed
Type Return	Quote

VIAMEDclean

Goods In Only
 Decontamination certificate provided by customer

Goods Out Only
 Cleaned by Viamed before returning to customer

Cleaned by Viamed, if no declaration certificate from customer

Signed: _____
 Date: _____

Signed: _____
 Date: _____

Notes Has a faulty MySign O (serial number 100107). It does not charge even when a different plug is used.

06/Jan/2026 Kate Griffiths
 update, customer will send it in, asked for cert of decont/no batteries/SRS on all paperwork. Sophie had asked for different s/n, but customer says it is as stated

08/Jan/2026 Robert Connor
 Received 1 x MySign O s/n 1000107, no accessories.

16/Jan/2026 Catherine Spence
 requesting RMA on Box1063

Ready For quote

CCMEON 10-2-26

Repair Complete Signed _____

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38511	Oxygen Monitor	0111276	100107	N

0140000 x 1 1/2 @ ⁹⁵ £90 per hr

S/N, SRS, SRN

UPS x 1 £12.