

Service Repair Sheet 69348

Contact Name Sarah Fairley  
Company/ Hospital Name Russells Hall Hospital  
Department Medical Engineering Department  
Position Secretary/Administrator  
Direct Phone 01384456111  
General Phone 01384456111  
Opera Account 00001465  
Email sarah.fairley2@nhs.net  
Order Number 500006412-01  
Date Received 05/Feb/2026  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38528	1410000	PR03485A16	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.

---