

Service Repair Sheet 69346
Contact Name Derek Ng
Company/ Hospital Name Tronda Electronics Limited
Position Service Engineer
Direct Phone 85226482822
General Phone 85226482822
Opera Account CID35882
Email derek@tronda.com.hk
Order Number TP26 265064
Date Received 16/Feb/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN38541	1410000	PR03744A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
